
Midmark M11 - v1.0.0
Total Cycles: 4794

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2024 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 177.3 F 1.9
4:00 181.3 F 5.2
6:00 199.4 F 9.2
8:00 214.0 F 13.3
10:00 226.0 F 6.2
12:00 235.9 F 10.4
14:00 233.1 F 8.4
16:00 242.8 F 11.9
18:00 239.6 F 10.0
20:00 251.8 F 15.9
22:00 262.1 F 21.9
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.3 F 28.8
2:00 272.1 F 28.5
3:00 271.5 F 28.1
4:00 271.1 F 28.1

Min 271.0 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 221.4 F 0.0
10:00 216.1 F 0.0
15:00 191.1 F 0.0
20:00 187.6 F 0.0
25:00 187.0 F 0.0
30:00 185.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

08/08/2024 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____