

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2206

Name: RITTER2  
Sterilizer ID: V2313228  
04/03/2023 10:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.8 F 0.0  
2:00 174.2 F 2.2  
4:00 191.3 F 6.6  
6:00 210.8 F 12.2  
8:00 226.4 F 7.9  
10:00 239.7 F 12.6  
12:00 240.7 F 11.9  
14:00 240.4 F 11.0  
16:00 253.9 F 18.0  
18:00 265.8 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.5 F 29.6  
2:00 271.5 F 29.2  
3:00 272.5 F 29.6  
4:00 271.5 F 28.9  
  
Min 271.4 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.2 F 0.0  
5:00 225.5 F 0.0  
10:00 222.7 F 0.0  
15:00 223.7 F 0.0  
20:00 230.2 F 0.0  
25:00 227.3 F 0.0  
30:00 229.5 F 0.0  
  
Drying Complete

FILLING: 1:51  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

04/03/2023 11:29 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_