

-----  
Midmark M11 - v1.0.0  
Total Cycles: 799

Name: RITTER 1  
Sterilizer ID: V2178198  
04/22/2021 09:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.8 F 0.0  
2:00 135.2 F 0.7  
4:00 148.8 F 2.7  
6:00 172.9 F 5.9  
8:00 194.9 F 10.0  
10:00 216.5 F 6.2  
12:00 225.6 F 8.6  
14:00 237.8 F 13.6  
16:00 238.0 F 10.6  
18:00 234.8 F 8.4  
20:00 249.5 F 15.1  
22:00 261.6 F 21.9  
24:00 272.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.5 F 29.3  
2:00 272.1 F 28.7  
3:00 271.2 F 28.4  
4:00 272.4 F 29.3

Min 271.1 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 229.4 F 0.0  
10:00 227.8 F 0.0  
15:00 227.9 F 0.0  
20:00 223.8 F 0.0  
Dry Cycle Aborted

FILLING: 1:00  
HEATING: 24:29  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 23:19  
TOTAL CYCLE: 00:56:04

04/22/2021 10:47 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_