
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005169
14:22 28/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:18
136.4 C 220kPa 4:18
136.2 C 219kPa 4:48
136.4 C 224kPa 5:18
136.0 C 222kPa 5:48
136.3 C 224kPa 6:18
136.2 C 221kPa 6:48
136.4 C 224kPa 7:18
136.4 C 223kPa 7:48
136.4 C 223kPa 7:48
VENTING 7:48
AIR DRYING 8:56
CYCLE COMPLETE 10:10

Digital Signature #
509FE4D4E68E32BE
