
Midmark M11 - v1.0.0
Total Cycles: 3643

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2023 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 175.1 F 3.0
4:00 187.0 F 6.5
6:00 218.2 F 11.0
8:00 220.2 F 4.8
10:00 231.2 F 9.5
12:00 237.6 F 10.1
14:00 240.8 F 11.3
16:00 237.9 F 9.6
18:00 251.1 F 15.7
20:00 261.7 F 21.8
22:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.1 F 28.6
2:00 272.1 F 28.9
3:00 272.6 F 29.2
4:00 271.9 F 29.0

Min 271.5 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 229.2 F 0.0
10:00 221.0 F 0.0
15:00 229.2 F 0.0
20:00 224.1 F 0.0
25:00 226.8 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

08/14/2023 12:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____