
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004396
13:31 20/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:14
PRESSURIZING 1:19
STERILIZING 2:09
136.0 C 220kPa 2:09
136.4 C 222kPa 2:39
136.2 C 223kPa 3:09
136.3 C 222kPa 3:39
136.1 C 224kPa 4:09
136.1 C 223kPa 4:39
136.4 C 222kPa 5:09
136.4 C 224kPa 5:39
136.4 C 224kPa 5:39
VENTING 5:39
AIR DRYING 6:27
CYCLE COMPLETE 6:43

Digital Signature #
C516027619828ACE
