

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3911

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2023 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
2:00 160.4 F 3.0  
4:00 177.1 F 6.8  
6:00 197.5 F 11.5  
8:00 219.5 F 5.2  
10:00 232.8 F 10.1  
12:00 233.7 F 7.8  
14:00 242.1 F 12.2  
16:00 240.0 F 10.4  
18:00 253.2 F 16.9  
20:00 264.7 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.7 F 28.9  
2:00 271.5 F 28.7  
3:00 272.9 F 29.1  
4:00 271.6 F 28.8

Min 270.9 F 28.3  
Max 273.2 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 255.0 F 0.0  
10:00 236.5 F 0.0  
15:00 224.2 F 0.0  
20:00 228.0 F 0.0  
25:00 230.0 F 0.0  
30:00 237.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 21:53  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:00:24  
CYCLE COMPLETE

10/17/2023 12:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_