
Midmark M11 - v1.0.0
Total Cycles: 3968

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2023 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 151.4 F 2.2
4:00 158.4 F 5.3
6:00 194.3 F 9.0
8:00 214.7 F 13.5
10:00 218.4 F 7.1
12:00 233.7 F 11.6
14:00 232.2 F 7.9
16:00 243.8 F 13.5
18:00 243.0 F 12.0
20:00 254.7 F 18.0
22:00 265.5 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.7 F 29.3
2:00 271.8 F 28.7
3:00 271.4 F 28.8
4:00 272.7 F 29.3

Min 271.2 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 214.3 F 0.0
10:00 230.1 F 0.0
15:00 232.1 F 0.0
20:00 228.0 F 0.0
25:00 229.9 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

11/01/2023 04:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____