
Midmark M11 - v1.0.5
Total Cycles: 40

Name: RITTER2
Sterilizer ID: V2313228
05/20/2021 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 170.8 F 3.1
4:00 189.3 F 7.2
6:00 208.2 F 11.8
8:00 218.0 F 6.2
10:00 231.1 F 11.2
12:00 233.9 F 9.0
14:00 241.1 F 10.9
16:00 247.7 F 14.4
18:00 258.9 F 20.8
20:00 269.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.7
2:00 272.2 F 29.1
3:00 271.7 F 29.1
4:00 272.6 F 29.7

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 216.6 F 0.0
10:00 206.8 F 0.0
15:00 205.0 F 0.0
20:00 202.0 F 0.0
25:00 198.5 F 0.0
30:00 196.0 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

05/20/2021 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____