
Midmark M11 - v1.0.0
Total Cycles: 4280

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2024 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.0 F 0.0
2:00 148.6 F 2.5
4:00 157.3 F 6.2
6:00 208.9 F 10.6
8:00 217.4 F 4.6
10:00 225.2 F 10.2
12:00 229.5 F 7.1
14:00 242.5 F 12.6
16:00 240.6 F 10.7
18:00 253.6 F 17.2
20:00 264.9 F 24.0
22:00 270.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 271.7 F 28.7
2:00 272.8 F 29.1
3:00 272.3 F 29.3
4:00 271.4 F 28.7

Min 271.3 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 231.9 F 0.0
10:00 246.9 F 0.0
15:00 232.3 F 0.0
20:00 226.5 F 0.0
25:00 225.9 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

02/07/2024 02:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____