
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 004764
17:30 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:02
STERILIZING 4:04
136.4 C 222kPa 4:04
136.4 C 225kPa 4:34
136.4 C 225kPa 5:04
136.3 C 225kPa 5:34
136.4 C 224kPa 6:04
136.3 C 225kPa 6:34
136.2 C 223kPa 7:04
136.3 C 224kPa 7:34
136.3 C 224kPa 7:34
VENTING 7:34
AIR DRYING 8:28
CYCLE COMPLETE 10:14

Digital Signature #
8B9A2E9FCB91FE9C
