
Midmark M11 - v1.0.5
Total Cycles: 1253

Name: RITTER2
Sterilizer ID: V2313228
07/05/2022 06:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 91.4 F 0.0
4:00 85.3 F 0.0
6:00 103.0 F 0.0
8:00 143.5 F 0.4
10:00 170.4 F 2.9
12:00 201.3 F 6.0
14:00 199.4 F 9.7
16:00 217.1 F 13.4
18:00 224.5 F 7.9
20:00 239.0 F 12.2
22:00 234.5 F 9.0
24:00 241.1 F 11.0
26:00 244.5 F 12.8
28:00 255.9 F 18.7
30:00 265.9 F 25.1
32:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 271.8 F 29.7
2:00 272.2 F 29.8
3:00 272.7 F 29.8
4:00 272.3 F 29.7

Min 271.2 F 29.3
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 227.7 F 0.0
10:00 219.6 F 0.0
15:00 185.0 F 0.0
20:00 199.4 F 0.0
25:00 193.7 F 0.0
30:00 209.6 F 0.0

Drying Complete

FILLING: 1:02
HEATING: 32:24
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:10:32
CYCLE COMPLETE

07/05/2022 07:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____