
Midmark M11 - v1.0.0
Total Cycles: 1715

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2022 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 177.1 F 3.0
4:00 190.7 F 7.4
6:00 217.5 F 13.3
8:00 225.9 F 8.2
10:00 238.9 F 12.9
12:00 239.7 F 11.0
14:00 237.6 F 9.3
16:00 252.0 F 16.3
18:00 263.9 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.3
3:00 272.6 F 29.0
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 221.5 F 0.0
10:00 227.2 F 0.0
15:00 232.3 F 0.0
20:00 227.1 F 0.0
25:00 227.1 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

03/14/2022 02:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____