

-----  
Midmark M11 - v1.0.5  
Total Cycles: 692

Name: RITTER2  
Sterilizer ID: V2313228  
01/13/2022 07:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.5 F 0.0  
2:00 77.2 F 0.0  
4:00 80.1 F 0.0  
6:00 98.9 F 0.0  
8:00 132.5 F 1.5  
10:00 159.2 F 4.0  
12:00 179.0 F 7.1  
14:00 202.6 F 10.8  
16:00 217.3 F 5.0  
18:00 226.8 F 9.0  
20:00 236.6 F 12.4  
22:00 238.9 F 10.9  
24:00 236.0 F 9.1  
26:00 249.7 F 15.5  
28:00 260.8 F 21.8  
30:00 271.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.1 F 29.1  
2:00 272.4 F 29.6  
3:00 271.4 F 29.2  
4:00 271.9 F 29.1  
  
Min 271.2 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
5:00 226.0 F 0.0  
10:00 217.0 F 0.0  
15:00 209.5 F 0.0  
20:00 207.0 F 0.0  
25:00 230.7 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 30:41  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:51  
CYCLE COMPLETE

01/13/2022 09:17 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02