
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001519
12:51 19/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:15
STERILIZING 4:22
136.0 C 221kPa 4:22
136.4 C 226kPa 4:52
136.4 C 226kPa 5:22
136.3 C 224kPa 5:52
136.4 C 225kPa 6:22
136.1 C 222kPa 6:52
136.4 C 224kPa 7:22
136.4 C 225kPa 7:52
136.4 C 225kPa 7:52
VENTING 7:52
AIR DRYING 8:46
CYCLE COMPLETE 8:53

Digital Signature #
DA1800BF6061685B
