

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1016

Name: RITTER2  
Sterilizer ID: V2313228  
04/21/2022 08:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.5 F 0.0  
2:00 163.9 F 0.1  
4:00 171.1 F 0.6  
6:00 194.7 F 4.6  
8:00 215.8 F 9.4  
10:00 229.5 F 9.0  
12:00 235.8 F 10.4  
14:00 235.6 F 8.4  
16:00 244.1 F 12.1  
18:00 245.9 F 13.3  
20:00 258.5 F 20.3  
22:00 270.0 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.7 F 29.3  
2:00 271.8 F 28.5  
3:00 272.1 F 29.3  
4:00 272.6 F 29.1

Min 271.3 F 28.4  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
5:00 233.7 F 0.0  
10:00 216.2 F 0.0  
15:00 220.8 F 0.0  
20:00 219.2 F 0.0  
25:00 224.6 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 22:57  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:14  
CYCLE COMPLETE

04/21/2022 09:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_