

-----  
Midmark M11 - v1.0.5  
Total Cycles: 321

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2021 12:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.6 F 0.0  
2:00 157.1 F 3.1  
4:00 169.0 F 7.2  
6:00 193.4 F 12.1  
8:00 209.6 F 5.8  
10:00 226.1 F 10.7  
12:00 231.6 F 7.9  
14:00 241.0 F 12.4  
16:00 244.0 F 13.0  
18:00 256.0 F 19.5  
20:00 267.0 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.7  
1:00 272.2 F 30.4  
2:00 271.7 F 29.6  
3:00 272.2 F 30.4  
4:00 272.0 F 29.8

Min 271.3 F 29.5  
Max 272.6 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 235.4 F 0.0  
10:00 219.1 F 0.0  
15:00 204.8 F 0.0  
20:00 196.9 F 0.0  
25:00 188.0 F 0.0  
30:00 185.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

08/30/2021 01:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_