
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000050
12:19 16/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:33
STERILIZING 4:40
136.0 C 220kPa 4:40
136.2 C 224kPa 5:10
136.1 C 220kPa 5:40
136.3 C 224kPa 6:10
136.4 C 224kPa 6:40
136.3 C 224kPa 7:10
136.3 C 220kPa 7:40
136.1 C 224kPa 8:10
136.1 C 224kPa 8:10
VENTING 8:10
AIR DRYING 9:02
CYCLE COMPLETE 9:10

Digital Signature #
935244C6F11604A2
