
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002463
11:51 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:16
136.0 C 221kPa 4:16
136.2 C 224kPa 4:46
136.4 C 226kPa 5:16
136.4 C 223kPa 5:46
136.4 C 227kPa 6:16
136.4 C 226kPa 6:46
136.4 C 226kPa 7:16
136.4 C 226kPa 7:46
136.4 C 226kPa 7:46
VENTING 7:46
AIR DRYING 8:41
CYCLE COMPLETE 8:53

Digital Signature #
D59184E981BF4FD6
