
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002418
10:36 13/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:42
STERILIZING 3:49
136.4 C 224kPa 3:49
136.3 C 225kPa 4:19
136.4 C 226kPa 4:49
136.4 C 226kPa 5:19
136.4 C 226kPa 5:49
136.4 C 225kPa 6:19
136.2 C 224kPa 6:49
136.4 C 226kPa 7:19
136.4 C 226kPa 7:19
VENTING 7:19
AIR DRYING 8:13
CYCLE COMPLETE 31:28

Digital Signature #
6D2145FB35EDA242
