
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000173
12:21 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:03
PRESSURIZING 4:05
STERILIZING 5:32
136.4 C 224kPa 5:32
136.4 C 225kPa 6:02
136.4 C 226kPa 6:32
136.3 C 225kPa 7:02
136.4 C 225kPa 7:32
136.4 C 226kPa 8:02
136.4 C 225kPa 8:32
136.3 C 225kPa 9:02
136.3 C 225kPa 9:02
VENTING 9:02
AIR DRYING 10:04
CYCLE COMPLETE 25:04

Digital Signature #
CA94BAB41C1CF417
