

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5371

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2025 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.0 F 0.0  
2:00 172.7 F 3.6  
4:00 183.6 F 7.9  
6:00 211.0 F 12.5  
8:00 221.0 F 6.4  
10:00 231.7 F 11.2  
12:00 232.9 F 7.9  
14:00 245.1 F 12.8  
16:00 244.0 F 12.4  
18:00 255.6 F 18.8  
20:00 267.1 F 25.5  
22:00 273.0 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.6 F 29.4  
2:00 271.8 F 29.2  
3:00 272.6 F 29.4  
4:00 271.7 F 29.1

Min 271.7 F 28.9  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 231.8 F 0.0  
10:00 218.7 F 0.0  
15:00 216.9 F 0.0  
20:00 205.3 F 0.0  
25:00 205.1 F 0.0  
30:00 200.0 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 22:20  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:02:13  
CYCLE COMPLETE

02/26/2025 05:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_