
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001507
9:11 18/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:04
STERILIZING 4:12
136.1 C 220kPa 4:12
136.3 C 223kPa 4:42
136.0 C 220kPa 5:12
136.3 C 223kPa 5:42
136.3 C 221kPa 6:12
136.2 C 224kPa 6:42
136.4 C 224kPa 7:12
136.4 C 222kPa 7:42
136.4 C 222kPa 7:42
VENTING 7:42
AIR DRYING 8:33
CYCLE COMPLETE 9:00

Digital Signature #
7110ABE27B401E28
