
Midmark M11 - v1.0.0
Total Cycles: 5093

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2024 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.4 F 0.0
2:00 138.3 F 0.2
4:00 162.1 F 0.6
6:00 189.3 F 1.0
8:00 217.1 F 4.0
10:00 221.9 F 7.8
12:00 229.9 F 11.8
14:00 231.3 F 7.3
16:00 241.6 F 12.5
18:00 240.2 F 10.3
20:00 236.9 F 8.6
22:00 250.5 F 15.1
24:00 261.6 F 21.4
26:00 271.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 271.8 F 28.4
2:00 272.8 F 28.9
3:00 272.1 F 28.3
4:00 272.2 F 28.7

Min 271.0 F 27.7
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 230.2 F 0.0
10:00 226.8 F 0.0
15:00 236.1 F 0.0
20:00 230.5 F 0.0
25:00 227.1 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:30
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:14
CYCLE COMPLETE

11/14/2024 05:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____