

-----  
Midmark M11 - v1.0.0  
Total Cycles: 805

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2021 07:56 AM

BEGIN PACKS CYCLE  
Temp: 250 Degrees F  
Time: 30 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.6 F 0.0  
2:00 138.4 F 1.0  
4:00 149.5 F 3.6  
6:00 176.6 F 6.7  
8:00 204.0 F 2.1  
10:00 218.0 F 6.2  
12:00 223.9 F 3.9  
14:00 228.6 F 5.4  
16:00 235.6 F 8.1  
18:00 247.6 F 13.6

STERILIZING  
mm:ss Degrees PSI  
0:00 251.6 F 15.6  
1:00 252.5 F 16.0  
2:00 251.8 F 15.9  
3:00 252.4 F 15.9  
4:00 251.6 F 15.7  
5:00 252.6 F 16.1  
6:00 252.0 F 15.7  
7:00 252.0 F 16.0  
8:00 252.5 F 16.1  
9:00 252.1 F 15.8  
10:00 251.5 F 15.6  
11:00 252.6 F 16.3  
12:00 252.4 F 16.1  
13:00 252.1 F 15.8  
14:00 251.7 F 15.6  
15:00 251.8 F 15.9  
16:00 252.7 F 16.3  
17:00 252.4 F 16.1  
18:00 252.1 F 15.9  
19:00 251.8 F 15.6  
20:00 251.6 F 15.8  
21:00 252.6 F 16.3  
22:00 252.6 F 16.2  
23:00 252.3 F 16.0  
24:00 252.1 F 15.8  
25:00 251.8 F 15.6  
26:00 251.5 F 15.8  
27:00 252.6 F 16.3  
28:00 252.8 F 16.2  
29:00 252.6 F 16.1  
30:00 252.3 F 15.9

Min 251.3 F 15.5  
Max 252.8 F 16.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 180.7 F 0.0  
10:00 187.1 F 0.0  
15:00 208.1 F 0.0  
20:00 195.7 F 0.0  
25:00 191.7 F 0.0  
30:00 182.7 F 0.0  
Drying Complete

FILLING: 0:59

HEATING: 19:19  
STERILIZING: 30:00  
VENTING: 2:54  
DRYING: 30:00  
TOTAL CYCLE: 01:23:13  
CYCLE COMPLETE

04/26/2021 09:19 AM

ACCEPTED:

REJECTED:

Reason: -----  
Operator: -----