
Midmark M11 - v1.0.5
Total Cycles: 1108

Name: RITTER2
Sterilizer ID: V2313228
05/18/2022 10:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 167.1 F 2.0
4:00 183.0 F 5.6
6:00 214.7 F 10.3
8:00 223.7 F 6.3
10:00 231.6 F 10.4
12:00 231.5 F 7.8
14:00 242.5 F 13.1
16:00 241.8 F 11.7
18:00 253.9 F 18.0
20:00 264.6 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.4 F 30.0
2:00 271.8 F 29.3
3:00 271.6 F 29.7
4:00 272.3 F 29.9

Min 271.3 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 216.6 F 0.0
10:00 228.5 F 0.0
15:00 231.1 F 0.0
20:00 229.3 F 0.0
25:00 228.0 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:60
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

05/18/2022 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____