

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001291  
12:52 26/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:18  
STERILIZING 4:26  
136.3 C 223kPa 4:26  
CYCLE ABORTED 4:54  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 4:54  
-----

Digital Signature #  
9043EB24B383938C  
-----