
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003923
9:19 04/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:20
STERILIZING 4:31
136.4 C 222kPa 4:31
136.4 C 224kPa 5:01
136.3 C 223kPa 5:31
136.4 C 223kPa 6:01
136.3 C 222kPa 6:31
136.1 C 217kPa 7:01
136.1 C 221kPa 7:31
136.4 C 223kPa 8:01
136.4 C 223kPa 8:01
VENTING 8:02
AIR DRYING 9:02
CYCLE COMPLETE 9:10

Digital Signature #
19333FF7E036767E
