

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1687

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2022 11:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 178.1 F 0.5  
4:00 195.1 F 4.0  
6:00 213.9 F 9.2  
8:00 225.2 F 8.7  
10:00 235.6 F 11.6  
12:00 236.6 F 9.9  
14:00 236.7 F 8.7  
16:00 251.3 F 16.3  
18:00 263.7 F 23.6  
20:00 271.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.5 F 29.5  
2:00 271.6 F 29.0  
3:00 272.7 F 29.5  
4:00 271.6 F 28.8

Min 271.5 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 225.8 F 0.0  
10:00 240.4 F 0.0  
15:00 245.7 F 0.0  
20:00 229.1 F 0.0  
25:00 227.6 F 0.0  
30:00 221.2 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 20:01  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

11/07/2022 12:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_