
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001732
10:48 10/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:18
STERILIZING 3:22
136.5 C 224kPa 3:22
136.3 C 223kPa 3:52
136.0 C 223kPa 4:22
136.2 C 221kPa 4:52
136.2 C 222kPa 5:22
136.1 C 219kPa 5:52
136.1 C 224kPa 6:22
136.4 C 224kPa 6:52
136.4 C 224kPa 6:52
VENTING 6:52
AIR DRYING 7:44
CYCLE COMPLETE 7:59

Digital Signature #
AAE36AA689EB390C
