

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4326

Name: RITTER2  
Sterilizer ID: V2313228  
02/03/2025 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.8 F 0.0  
2:00 73.0 F 0.0  
4:00 90.8 F 0.3  
6:00 110.3 F 2.9  
8:00 145.3 F 5.9  
10:00 168.9 F 9.1  
12:00 188.0 F 12.8  
14:00 211.6 F 3.7  
16:00 220.8 F 7.1  
18:00 232.8 F 11.0  
20:00 229.3 F 7.2  
22:00 240.2 F 11.6  
24:00 236.6 F 9.4  
26:00 249.5 F 15.4  
28:00 260.3 F 21.6  
30:00 270.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.5 F 29.3  
2:00 272.2 F 29.6  
3:00 272.0 F 28.9  
4:00 272.7 F 29.7

Min 271.7 F 28.8  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 192.7 F 0.0  
10:00 208.6 F 0.0  
15:00 200.7 F 0.0  
20:00 200.1 F 0.0  
25:00 199.2 F 0.0  
30:00 199.3 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:48  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:09:06  
CYCLE COMPLETE

02/03/2025 11:06 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_