
Midmark M11 - v1.0.0
Total Cycles: 1298

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2021 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.4 F 0.0
2:00 120.0 F 0.0
4:00 168.9 F 0.0
6:00 181.6 F 0.1
8:00 188.4 F 0.2
10:00 203.9 F 2.6
12:00 224.3 F 5.8
14:00 231.3 F 9.3
16:00 240.2 F 13.2
18:00 237.2 F 9.1
20:00 243.6 F 11.5
22:00 243.4 F 11.7
24:00 239.7 F 10.0
26:00 252.6 F 16.2
28:00 263.4 F 22.5
30:00 272.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 272.2 F 28.6
2:00 271.9 F 28.0
3:00 272.0 F 28.5
4:00 272.3 F 28.3

Min 271.3 F 27.7
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 228.1 F 0.0
10:00 224.7 F 0.0
15:00 219.6 F 0.0
20:00 213.7 F 0.0
25:00 208.9 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:08:16
CYCLE COMPLETE

10/25/2021 03:32 PM
ACCEPTED: X
REJECTED:
Operator: U01