

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2371

Name: RITTER2  
Sterilizer ID: V2313228  
05/16/2023 07:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.5 F 0.0  
2:00 76.9 F 0.0  
4:00 81.2 F 0.0  
6:00 104.2 F 0.3  
8:00 139.4 F 0.4  
10:00 174.1 F 1.6  
12:00 193.7 F 4.8  
14:00 214.1 F 8.4  
16:00 224.2 F 12.5  
18:00 231.0 F 7.9  
20:00 240.6 F 12.7  
22:00 238.9 F 10.3  
24:00 235.3 F 8.6  
26:00 249.7 F 15.4  
28:00 261.2 F 21.9  
30:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.7  
1:00 272.1 F 29.5  
2:00 271.8 F 28.8  
3:00 271.6 F 29.4  
4:00 273.4 F 30.3  
  
Min 271.0 F 28.6  
Max 273.4 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
5:00 202.7 F 0.0  
10:00 196.8 F 0.0  
15:00 195.0 F 0.0  
20:00 194.2 F 0.0  
25:00 193.0 F 0.0  
30:00 193.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:31  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:33  
CYCLE COMPLETE

05/16/2023 08:16 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_