
Midmark M11 - v1.0.5
Total Cycles: 332

Name: RITTER2
Sterilizer ID: V2313228
09/02/2021 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.0 F 0.0
2:00 95.2 F 0.5
4:00 108.6 F 2.3
6:00 140.3 F 5.3
8:00 168.5 F 9.4
10:00 192.5 F 9.5
12:00 210.4 F 6.5
14:00 224.5 F 10.6
16:00 231.5 F 7.8
18:00 239.6 F 11.9
20:00 237.6 F 10.1
22:00 250.3 F 16.2
24:00 261.2 F 22.5
26:00 271.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.9
1:00 272.4 F 30.1
2:00 272.2 F 29.8
3:00 272.4 F 30.0
4:00 272.7 F 30.3

Min 271.7 F 29.8
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 223.7 F 0.0
10:00 210.2 F 0.0
15:00 202.8 F 0.0
20:00 197.3 F 0.0
25:00 195.2 F 0.0
30:00 201.8 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 27:16
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:06:00
CYCLE COMPLETE

09/02/2021 02:40 PM
ACCEPTED: X
REJECTED:
Operator: U02