

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4967

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2024 01:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.8 F 0.0  
2:00 118.1 F 0.2  
4:00 146.8 F 1.0  
6:00 167.6 F 3.0  
8:00 189.3 F 6.4  
10:00 213.1 F 10.4  
12:00 222.0 F 6.9  
14:00 228.4 F 7.8  
16:00 237.5 F 12.3  
18:00 235.9 F 8.7  
20:00 242.2 F 10.9  
22:00 245.4 F 12.7  
24:00 257.1 F 18.9  
26:00 267.7 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.7 F 29.1  
2:00 272.3 F 28.5  
3:00 271.3 F 28.3  
4:00 272.5 F 29.1

Min 271.2 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.3 F 0.0  
5:00 208.6 F 0.0  
10:00 193.4 F 0.0  
15:00 187.9 F 0.0  
20:00 186.9 F 0.0  
25:00 186.9 F 0.0  
30:00 185.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 27:27  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:05:59  
CYCLE COMPLETE

10/04/2024 02:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_