

-----  
Midmark M11 - v1.0.5  
Total Cycles: 361

Name: RITTER2  
Sterilizer ID: V2313228  
09/14/2021 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.7 F 0.0  
2:00 100.8 F 0.8  
4:00 125.3 F 2.9  
6:00 149.0 F 5.9  
8:00 172.6 F 9.4  
10:00 190.8 F 13.6  
12:00 212.8 F 6.6  
14:00 223.9 F 10.7  
16:00 228.5 F 7.8  
18:00 238.7 F 12.1  
20:00 238.2 F 10.3  
22:00 250.1 F 16.1  
24:00 260.6 F 22.1  
26:00 270.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.2 F 30.2  
2:00 272.4 F 29.8  
3:00 271.7 F 29.3  
4:00 271.2 F 29.5

Min 271.0 F 29.2  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 218.2 F 0.0  
10:00 222.4 F 0.0  
15:00 216.8 F 0.0  
20:00 212.9 F 0.0  
25:00 215.2 F 0.0  
Dry Cycle Aborted

FILLING: 1:02  
HEATING: 26:56  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 27:41  
TOTAL CYCLE: 01:02:55

09/14/2021 12:37 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01