
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002365
10:52 05/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:33
STERILIZING 4:35
136.0 C 220kPa 4:35
136.4 C 223kPa 5:05
136.3 C 222kPa 5:35
136.3 C 222kPa 6:05
136.4 C 224kPa 6:35
136.3 C 220kPa 7:05
136.4 C 221kPa 7:35
136.3 C 221kPa 8:05
136.3 C 221kPa 8:05
VENTING 8:06
AIR DRYING 9:02
CYCLE COMPLETE 9:05

Digital Signature #
5529B079557BE31B
