

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1496

Name: RITTER 1  
Sterilizer ID: V2178198  
01/05/2022 11:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.2 F 0.0  
2:00 167.4 F 1.5  
4:00 175.6 F 4.9  
6:00 190.5 F 9.1  
8:00 209.7 F 9.3  
10:00 221.9 F 7.1  
12:00 233.5 F 11.9  
14:00 233.1 F 8.5  
16:00 241.4 F 11.2  
18:00 244.4 F 12.7  
20:00 256.4 F 19.1  
22:00 267.2 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 29.3  
2:00 271.6 F 28.5  
3:00 272.6 F 29.5  
4:00 272.2 F 28.9

Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 211.0 F 0.0  
10:00 229.0 F 0.0  
15:00 226.6 F 0.0  
20:00 220.7 F 0.0  
25:00 222.6 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 23:31  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:40  
CYCLE COMPLETE

01/05/2022 12:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_