
Midmark M11 - v1.0.0
Total Cycles: 5167

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2024 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 73.1 F 0.0
4:00 87.8 F 0.0
6:00 117.3 F 0.0
8:00 152.6 F 0.2
10:00 181.0 F 0.5
12:00 199.3 F 0.8
14:00 215.3 F 3.9
16:00 222.5 F 6.8
18:00 232.3 F 10.3
20:00 236.5 F 9.1
22:00 240.7 F 11.1
24:00 235.9 F 8.6
26:00 240.3 F 9.8
28:00 248.6 F 14.2
30:00 260.6 F 21.0
32:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.8 F 28.8
2:00 272.3 F 28.3
3:00 271.9 F 28.3
4:00 272.9 F 28.8

Min 271.7 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 186.8 F 0.0
10:00 192.8 F 0.0
15:00 220.7 F 0.0
20:00 234.9 F 0.0
25:00 230.2 F 0.0
30:00 229.0 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:45
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:50
CYCLE COMPLETE

12/06/2024 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____