
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000056
12:45 19/01/2021
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 4:06
PRESSURIZING 17:10
STERILIZING 19:06
136.2 C 224kPa 19:06
136.0 C 223kPa 19:36
136.4 C 225kPa 20:06
136.4 C 228kPa 20:36
136.4 C 226kPa 21:06
136.4 C 227kPa 21:36
136.4 C 227kPa 22:06
136.2 C 225kPa 22:36
136.4 C 224kPa 23:06
136.0 C 224kPa 23:36
136.4 C 227kPa 24:06
136.2 C 225kPa 24:36
136.3 C 226kPa 25:06
VENTING 25:06
AIR DRYING 26:30
CYCLE COMPLETE 56:30

Digital Signature #
357693FC43201EE3
