

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1467

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2022 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.7 F 0.0  
2:00 117.8 F 0.0  
4:00 140.3 F 0.0  
6:00 169.8 F 2.7  
8:00 186.7 F 7.2  
10:00 211.6 F 11.8  
12:00 226.8 F 6.9  
14:00 233.9 F 11.9  
16:00 234.5 F 9.0  
18:00 241.0 F 10.5  
20:00 244.9 F 13.5  
22:00 256.6 F 19.8  
24:00 267.2 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 272.5 F 30.3  
2:00 272.4 F 29.9  
3:00 271.6 F 29.3  
4:00 271.9 F 29.8

Min 271.4 F 29.3  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 188.4 F 0.0  
10:00 197.1 F 0.0  
15:00 200.9 F 0.0  
20:00 205.0 F 0.0  
25:00 218.9 F 0.0  
30:00 223.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:32  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

09/08/2022 09:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_