
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001534
16:01 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:32
STERILIZING 3:36
136.0 C 220kPa 3:36
136.4 C 226kPa 4:06
136.4 C 226kPa 4:36
136.4 C 226kPa 5:06
136.4 C 226kPa 5:36
136.1 C 223kPa 6:06
136.4 C 226kPa 6:36
136.3 C 226kPa 7:06
136.3 C 226kPa 7:06
VENTING 7:06
AIR DRYING 8:03
CYCLE COMPLETE 8:19

Digital Signature #
2627EEF5C729C8E9
