

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1561

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2022 12:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
2:00 196.8 F 1.3  
4:00 212.4 F 6.7  
6:00 230.2 F 12.1  
8:00 232.2 F 9.0  
10:00 239.1 F 10.1  
12:00 245.4 F 12.9  
14:00 246.2 F 13.4  
16:00 258.9 F 20.5  
18:00 270.2 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.4 F 29.3  
2:00 271.6 F 28.6  
3:00 272.2 F 29.5  
4:00 272.3 F 29.1  
  
Min 271.2 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 226.1 F 0.0  
10:00 236.8 F 0.0  
15:00 229.3 F 0.0  
20:00 227.0 F 0.0  
25:00 226.3 F 0.0  
30:00 224.4 F 0.0  
  
Drying Complete

FILLING: 1:47  
HEATING: 18:53  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:48  
CYCLE COMPLETE

10/04/2022 01:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_