
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001510
13:03 10/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:05
PRESSURIZING 5:36
STERILIZING 6:58
136.0 C 220kPa 6:58
136.1 C 219kPa 7:28
136.2 C 221kPa 7:58
136.4 C 221kPa 8:28
136.5 C 222kPa 8:58
136.4 C 222kPa 9:28
136.4 C 224kPa 9:58
136.5 C 221kPa 10:28
VENTING 10:28
AIR DRYING 11:32
CYCLE COMPLETE 41:32

Digital Signature #
79EAD534E191EC6D
