
Midmark M11 - v1.0.5
Total Cycles: 3015

Name: RITTER2
Sterilizer ID: V2313228
10/16/2023 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.9 F 0.0
2:00 124.8 F 1.0
4:00 143.8 F 3.7
6:00 176.3 F 8.2
8:00 206.2 F 13.3
10:00 217.9 F 7.5
12:00 230.1 F 12.5
14:00 233.0 F 9.3
16:00 236.7 F 9.5
18:00 246.2 F 14.3
20:00 258.9 F 21.2
22:00 270.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.4 F 29.9
2:00 271.2 F 29.5
3:00 272.5 F 30.0
4:00 271.3 F 29.3

Min 271.1 F 29.2
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 209.0 F 0.0
10:00 212.8 F 0.0
15:00 214.1 F 0.0
20:00 215.2 F 0.0
25:00 216.5 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

10/16/2023 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____