
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004674
15:19 07/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 3:59
136.0 C 220kPa 3:59
136.4 C 225kPa 4:29
136.4 C 224kPa 4:59
136.2 C 222kPa 5:29
136.4 C 224kPa 5:59
136.4 C 223kPa 6:29
136.4 C 223kPa 6:59
136.1 C 223kPa 7:29
136.1 C 223kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 8:55

Digital Signature #
74F41D3F9D20A0C4
