
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000005
12:12 30/11/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:49
PRESSURIZING 1:03
STERILIZING 2:06
136.4 C 228kPa 2:06
136.4 C 226kPa 2:36
136.3 C 223kPa 3:06
136.3 C 223kPa 3:36
136.3 C 224kPa 4:06
136.4 C 219kPa 4:36
136.4 C 224kPa 5:06
136.3 C 222kPa 5:36
136.3 C 222kPa 5:36
VENTING 5:36
AIR DRYING 6:24
CYCLE COMPLETE 36:24

Digital Signature #
E1DB0F3CCEB9C5A2
