

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006607  
9:19 14/11/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:27  
STERILIZING 3:30  
136.4 C 223kPa 3:30  
136.1 C 222kPa 4:00  
136.3 C 222kPa 4:30  
136.3 C 221kPa 5:00  
136.0 C 222kPa 5:30  
136.4 C 223kPa 6:00  
136.3 C 223kPa 6:30  
136.4 C 223kPa 7:00  
136.4 C 223kPa 7:00  
VENTING 7:00  
AIR DRYING 7:50  
CYCLE COMPLETE 7:54

-----

Digital Signature #  
98941FC8DB612863  
-----