

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4701

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2024 12:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.4 F 0.0  
2:00 179.8 F 3.1  
4:00 190.2 F 7.1  
6:00 207.2 F 11.2  
8:00 219.7 F 4.7  
10:00 229.3 F 9.2  
12:00 237.3 F 11.3  
14:00 239.3 F 11.2  
16:00 236.8 F 9.3  
18:00 249.7 F 15.4  
20:00 260.5 F 21.4  
22:00 270.4 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.2  
2:00 271.3 F 28.5  
3:00 272.2 F 29.4  
4:00 272.2 F 29.0

Min 271.1 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 207.9 F 0.0  
10:00 212.4 F 0.0  
15:00 212.3 F 0.0  
20:00 204.2 F 0.0  
25:00 204.6 F 0.0  
30:00 206.4 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 22:53  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:02:17  
CYCLE COMPLETE

07/05/2024 01:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_