

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002510  
11:27 11/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:12  
STERILIZING 4:14  
136.4 C 225kPa 4:14  
136.4 C 224kPa 4:44  
136.1 C 223kPa 5:14  
136.4 C 225kPa 5:44  
136.4 C 226kPa 6:14  
136.3 C 224kPa 6:44  
136.4 C 223kPa 7:14  
136.4 C 223kPa 7:44  
136.4 C 223kPa 7:44  
VENTING 7:44  
AIR DRYING 8:37  
CYCLE COMPLETE 9:27  
-----

Digital Signature #  
864D49F1D29A484F  
-----